



Ref. Certif. No.

DK-149139-M1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

Sinpro Electronics Co Ltd
No. 5 Yuanxi St 10th Neighborhood
Dazhou Village Pingtung City Pingtung County 90093
Taiwan

Name and address of the manufacturer

Sinpro Electronics Co Ltd
No. 5 Yuanxi St 10th Neighborhood
Dazhou Village Pingtung City Pingtung County 90093
Taiwan

Name and address of the factory

Sinpro Electronics Co Ltd
No. 5 Yuanxi St 10th Neighborhood
Dazhou Village Pingtung City Pingtung County 90093
Taiwan

Note: When more than one factory, please report on page 2

☒ [Additional Information on page 2](#)

Ratings and principal characteristics

Input: 100-240 Vac, 47-63 Hz, 1.62-0.72 A,
See Test Report for output ratings.

Trademark / Brand (if any)

Sinpro

SINPRO SINPRO

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

MPU64-102, MPU64-103, MPU64-104, MPU64-105, MPU64-106,
MPU64-107, MPU64-108GW, MPU64-109, MPU64-110, MPU64-108

Additional information (if necessary may also be reported on page 2)

The report was revised to include technical modifications.
The risk management requirements of the standard were not addressed
National Differences: EU Group Differences, CA, JP, US
☒ [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012,
IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E230351-4790754494.27-1-Amendment 1 issued on 2024-08-19

This CB Test Certificate is issued by the National Certification Body



☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-08-21

Original Issue Date: 2024-01-04

Signature:

Thomas Wilson



Ref. Certif. No.

DK-149139-M1-UL

Factory(ies):

Agelong Technology Co Ltd
Floor 2, No.1 Workshop Building, Shengzuozhi Sci.&Tech. Industry Park
On the north of Xinyu Rd, Xiang Shan Community
XinQiao Subdistrict, Bao'an District Shenzhen, Guangdong, 518104
China

Additionally evaluated to:

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021

Summary of Modifications:

- This test report has been amended, due to the supplementary Temperature test measurement on Appliance Inlet outside (near the Line pin), and restoration of the Appliance Inlet vendor (RONG FENG INDUSTRIAL CO LTD., Type SS-7B-1, SS-120)

Additional information (if necessary)



☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2024-08-21

Original Issue Date: 2024-01-04

Signature:

Thomas Wilson